

MIMETAS

Instruction manual for safe use and warranty of

OrganoFlow®

Package contents

- OrganoFlow®
- Green tray inlay
- Power adapter (230/110 V AC to 12 V DC) and wall socket connectors
- Instruction manual




This document contains the necessary instructions to install and safely operate the OrganoFlow®. Please carefully read this manual prior to setting up and operating the device.

Getting started

Please follow the instructions below to install the OrganoFlow®.

1. After unboxing, place the OrganoFlow® on a levelled surface.
2. Plug in the AC/DC power adapter and switch on the OrganoFlow® by using the ON/OFF switch located at the back of the instrument.
3. Press the confirm button to calibrate the OrganoFlow®.
4. Use the up and down buttons to choose between interval or continuous mode. In interval mode, the OrganoFlow® will move the platform to the selected rocking angle and will remain at this position for the selected interval duration. In continuous mode, the OrganoFlow® will move the platform over the selected interval while never being static.
5. Choose the rocking angle and the interval using the up and down buttons.
6. Press the confirm button to start rocking.
7. Before installation of the OrganoFlow® in an incubator, make sure to switch it off and sterilize it as described in the service and maintenance section.



Intended use

The OrganoFlow® is intended for research applications only and is specifically developed for use in combination with the OrganoPlate® platform only. Besides its use in general lab conditions, the OrganoFlow® is designed to function in CO₂ incubators at temperatures between 5 and 40 °C (41 - 104°F). For indoor use only. When not in use, store the OrganoFlow at room temperature in a dry and dark place protected from direct sunlight.

Service and maintenance

The OrganoFlow® is designed to provide long lasting operation with minimal maintenance. No technical user maintenance is required and no serviceable parts are inside the OrganoFlow®. We recommend positioning the OrganoFlow® in a clean environment compatible with sterile cell culture practices (BSL1). The OrganoFlow® and its cable can be safely cleaned on a regular basis with 70% ethanol. For a deeper sterilization, a multipurpose disinfectant solution such as Virkon™ S or bleach can be used. Use a soaked tissue to wet the surfaces of the OrganoFlow®, including the plastic cord of the power cable, and incubate for a maximum of 10 minutes. To minimize degradation of the surfaces of the OrganoFlow®, the aluminum body in particular, wipe off residues of Virkon™ S and bleach with a sterile paper towel soaked in 70% ethanol or distilled water. If necessary, the green rocking tray and the metal rocking platform can be removed by unlatching the small screw at the back of the platform and lifting the tray. Always make sure the power is disconnected from the OrganoFlow® before cleaning.

	OrganoFlow® S	OrganoFlow® L
Dimensions LxWxH (cm)	16x16x8	45x16x8
Weight (kg)	1.3	2.4
Power consumption (W)	1.7	4.7
Maximum load	4 OrganoPlates®	16 OrganoPlates®
Maximum load (kg)	0.4	1.6
Sliding Tray	No	Yes
IP rating housing	IP54	
Tilt angle	0° to +25°	
Modes of operation	Interval or continuous	
Tilting interval	5s - 999 min	
Operating conditions	Up to 40°C (104°F) and 95% RH (non-condensing)	

Safety instructions

WARNING! Switch off the power and disconnect the power supply before lifting or moving the OrganoFlow® for service or maintenance to avoid electrical shock.

WARNING! DO NOT use the OrganoFlow® in hazardous conditions or expose it to hazardous materials/chemicals for which the device was not intended.

DO NOT lift the OrganoFlow® by the rocking tray to avoid mechanical damage.

WARNING! By putting too much force or weight on the

OrganoFlow®, the motor will move through its defined motor steps, resulting in an offset of the rocker angle. Please recalibrate the OrganoFlow® by restarting the device (toggle power switch OFF and ON). If an offset in calibration occurs without the user's notice, the OrganoFlow® would rock asymmetrically (for example, a 1 degree offset would result in 8 degrees rocking to the left and 6 degrees to the right when using an angle setpoint of 7 degrees). Nonetheless, this angle offset does not influence your experiment, because the fluid flow rate in the OrganoPlate® is driven by the difference and not the absolute inclination between the media reservoirs.

DO NOT insert fingers into any pinch points.

DO NOT immerse the OrganoFlow® in liquid for cleaning.

DO NOT place the power adapter inside a CO₂ incubator.

DO NOT replace the supplied adapter by any other adapter. If necessary, please contact Mimetax for support.

DO NOT place multiple devices in close proximity to one another (main bodies touching). This could allow the metal trays to run into each other, potentially causing mechanical damage to your OrganoFlow®.

DO NOT place the OrganoFlow® in front of the temperature sensor of the incubator.

DO NOT spray any liquid directly into the cooling fan at the back while cleaning.

DO NOT expose the OrganoFlow® to temperatures above 40 degrees C (104° F).

DO NOT remove the back label and open the OrganoFlow® for mechanical or electrical modifications. When your device shows signs of mechanical or electrical modifications or damage not related to normal laboratory use, warranty is void.

DO NOT put more than 4 OrganoFlow® devices in a single incubator to prevent overheating of your cultures. Distribute them evenly inside the incubator.

DO NOT place tissue culture plates behind the OrganoFlow®. The air flow created by the inbuilt fan could negatively affect your culture.

DO NOT place the OrganoFlow® too close to the back wall of the incubator to make sure the airflow of the inbuilt fans is not impeded. Leave at least 3 cm (>1 inch) of space available.

Safety regulations & standards

Mimetas hereby declares under its sole responsibility that the construction of this product conforms in accordance with the following standards:

Safety standards:

- NEN-EN-IEC-61010-1
- Safety requirements for electrical equipment for measurement, control and laboratory use.

EMC standards:

- NEN-EN-IEC-61000-3-2/3-3
- NEN-EN-IEC-61326-1
- RoHS compliant

Associated guidelines:

- EMC directive 2014/30/EC
- RoHS directive 2015/863
- RoHS2 directive 2011/65/EU
- Low Voltage Directive 2006/95/EC
- CE approved
- FCC approved

Warranty

Mimetas warrants the OrganoFlow® against defects in materials and workmanship for 2 years. Within this time period, Mimetas will replace or repair, without charge to the original purchaser, any defective part. The warranty is void if the product is defective due to product accident, product modification, connection to an improper electrical supply, lack of proper maintenance, contamination, improper installation or misuse. The warranty shall also not apply to defects arising from fire, flood, lightning or other conditions unrelated to correct

operation of the product. Warranty can be covered by, at Mimetas' election, (1) refund of the original purchaser's purchase price for the product (2) repair of the product, or (3) replacement of the product or defective parts. Evidence of purchase by the original purchaser is required. Mimetas makes no other warranty, expressed or implied, with respect to its products.

MIMETAS makes no warranty regarding the merchantability of the products or their suitability or fitness for any particular purpose of use.

Mimetas shall not be liable for, indirect, special, incidental or consequential damages of any nature. Any recovery for any claim shall be limited to the original purchase price for the product. If the OrganoFlow® is used in a manner not specified, the protection may be impaired.

Terms & conditions

The purchase and delivery are subject to the General Terms and Conditions of Delivery, Purchase and Use of Mimetas B.V.

www.mimetas.com/en/company/terms-conditions/

Contact info

For Mimetas product service information or support, please contact us by any means:

+31 (0)85 888 31 61
support@mimetas.com